



Duckworth & Kent Ltd

Titanium Reusable Ophthalmic Instrument Manufacturer



+44 (0) 1462 893254



@ info@duckworth-and-kent.co.uk



www.duckworth-and-kent.com

Terence House
7 Marquis Business Centre
Royston Road, Baldock
Hertfordshire, SG7 6XL
United Kingdom



ISO9001:2015 and ISO13485:2016 Certified

EU DECLARATION OF CONFORMITY

Manufacturer: Duckworth & Kent Ltd
Terence House
7 Marquis Business Centre
Royston Road
Baldock, Hertfordshire
SG7 6XL, United Kingdom

Authorised Representative: Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands

Duckworth & Kent Ltd declares that the medical devices listed in Table 1 and described hereafter,

Duckworth & Kent range of Invasive Measuring Devices

are in conformity with the essential requirements and provisions of the Council Directive 93/42/EEC as amended by Directive 2007/47/EC.

In accordance with Annex V (Metrological aspects only) of the Council Directive 93/42/EEC as amended by Directive 2007/47/EC,

These devices are designated as Class I Measuring

Notified Body: SGS Belgium NV (NB 1639)
Noorderlaan 87
2514 Antwerpen
Belgium

Classification Route

These devices are all transient use, reusable, non-sterile, invasive medical devices.

The classification is determined under Annex IX, Rule 5.

This declaration is issued under the sole responsibility of Duckworth & Kent Ltd.

.....
Martin Lock
Head of Regulatory Affairs



Table 1. Medical Devices covered under this Declaration of Conformity.

Measuring Devies Im, Invasive			
REF	Device Title	Description	Intended Use
6-182-2	Lacrima Dilator	Lacrima dilator	Intended to dilate the Lacrima puncta and therefore determine the size of punctum plug to be fitted.

Table 2. Relevant standards and common specifications.

Applicable Standards	
Standard Number	Title
EN 1041:2008	Information supplied by the manufacturer of medical devices
EN ISO 10993-1:2009	Biological evaluation of medical devices - Part 1: Evaluation and testing within a risk management process (ISO 10993-1:2009) EN ISO 10993-1:2009/AC:2010
EN ISO 13485:2016	Medical devices - Quality management systems - Requirements for regulatory purposes (ISO 13485:2016) EN ISO 13485:2016/AC:2018
EN ISO 14971:2012	Medical Devices - Application of risk management to medical devices (ISO 14971:2007, Corrected version 2007-10-01)
ISO 15223-1:2021	Medical devices - Symbols to be used with information to be supplied by the manufacturer - Part 1: General requirements
EN ISO 17664:2017	Processing of health care products - Information to be provided by the medical device manufacturer for the processing of medical devices (ISO 17664:2017)
EN 62366-1:2015+A1:2020	Medical devices - Part 1: Application of usability engineering to medical devices
93/42/EEC	Council Directive concerning medical devices
2007/47/EC	Council Directive amending 93/42/EEC
MEDDEV 2.7/1 rev.4 June 2016	Clinical evaluation: A guide for manufacturers and notified bodies
MEDDEV 2.12/1 rev.8 Jan 2013	Guidelines on a medical device's vigilance system



Duckworth & Kent Ltd

Titanium Reusable Ophthalmic Instrument Manufacturer



+44 (0) 1462 893254



Info@duckworth-and-kent.co.uk



www.duckworth-and-kent.com

Terence House
7 Marquis Business Centre
Royston Road, Baldock
Hertfordshire, SG7 6XL
United Kingdom

ISO13485:2016 Certified



ADDENDUM TO THE ORIGINAL DECLARATION OF CONFORMITY

Per 31 January 2023, the address of our EU Authorised Representative as listed on the original Declaration of Conformity has changed

Old Address

Authorised Representative:

Emergo Europe
Prinsessegracht 20
2514 AP The Hague
The Netherlands

New Address

Authorised Representative:

Emergo Europe
Westervoortsedijk 60
6827 AT Arnhem
The Netherlands

Martin Lock
Head of Regulatory Affairs

25/01/23

Date